



---

*Box R Water Analysis Laboratory*

567 NW Second Street  
Prineville, Oregon 97754  
541-447-4911

Mr. Wade Peasley

Aug 16, 2016

c/o Harney County School District #3

550 North Court Ave.

Burns, OR 97720

Sample Nbr: X023294


Dear Mr. Peasley,

Attached is a copy of your drinking water – Lead – test results, sampled on Aug 5, 2016, at L.J.H. School, Burns, OR. Your analysis was performed by Neilson Research Corp. in Medford, OR. Please do not hesitate to call Box R Water Analysis Laboratory with any questions you may have in regards to your water testing.

Thank you for using Box R Water Analysis Laboratory, we appreciate your business.

Sincerely,

Sherri K. Miyazaki – Box R Water Analysis Laboratory Director



# NEILSON RESEARCH CORPORATION

*Environmental Testing Laboratory*

8/16/2016

Sherri Miyazaki  
Box R Waterlab  
567 NW Second Street  
Prineville, OR 97754

TEL: (541) 447-4911

FAX (541) 447-4917

RE: X023294 L.J.H. School

Order No.: 1608474

Dear Sherri Miyazaki:

Neilson Research Corporation received 4 sample(s) on 8/9/2016 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,  
Neilson Research Corporation

Alec C Smith  
Project Manager

# Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

## Analysis Report

ORELAP 100016  
EPA OR00028

**CLIENT:** Box R Waterlab  
**Project:** X023294 L.J.H. School  
**Lab Order:** 1608474

**Date:** 16-Aug-16

## CASE NARRATIVE

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

# Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

## Analysis Report

ORELAP 100016  
EPA OR00028

**Box R Waterlab**  
567 NW Second Street  
Prineville, OR 97754

Lab Order: **1608474**  
NRC Sample ID: **1608474-01A**  
Collection Date: **8/5/2016 8:11:00 AM**  
Received Date: **8/9/2016 10:00:00 AM**  
Reported Date: **8/16/2016 7:33:44 AM**

### Sample Information:

**X023294 L.J.H. School**

Client Sample ID: X023294-01  
Collectors Name: Wade Peasley  
Sample Location: Gym DF  
Source: City Water

## ANALYTICAL RESULTS

| Analyses | Method    | NELAP      |         | Qual | MRL      | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|-----------|---------------|---------|
|          |           | Accredited | Result  |      |          |       |           |               |         |
| Lead     | EPA 200.8 | A          | 0.00374 |      | 0.000103 | mg/L  | 0.020 AL  | 8/10/2016     | OML     |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

# Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

## Analysis Report

ORELAP 100016  
EPA OR00028

**Box R Waterlab**  
567 NW Second Street  
Prineville, OR 97754

Lab Order: **1608474**  
NRC Sample ID: **1608474-02A**  
Collection Date: **8/5/2016 8:12:00 AM**  
Received Date: **8/9/2016 10:00:00 AM**  
Reported Date: **8/16/2016 7:33:44 AM**

### Sample Information:

**X023294 L.J.H. School**

Client Sample ID: X023294-02  
Collectors Name: Wade Peasley  
Sample Location: Office DF  
Source: City Water

## ANALYTICAL RESULTS

| Analyses | Method    | NELAP      |          | Qual | MRL      | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|----------|------|----------|-------|-----------|---------------|---------|
|          |           | Accredited | Result   |      |          |       |           |               |         |
| Lead     | EPA 200.8 | A          | 0.000675 |      | 0.000103 | mg/L  | 0.020 AL  | 8/10/2016     | OML     |

**Notes:** ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

# Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

## Analysis Report

ORELAP 100016  
EPA OR00028

**Box R Waterlab**  
567 NW Second Street  
Prineville, OR 97754

Lab Order: **1608474**  
NRC Sample ID: **1608474-03A**  
Collection Date: **8/5/2016 8:13:00 AM**  
Received Date: **8/9/2016 10:00:00 AM**  
Reported Date: **8/16/2016 7:33:44 AM**

### Sample Information:

**X023294 L.J.H. School**

Client Sample ID: X023294-03  
Collectors Name: Wade Peasley  
Sample Location: Sink in Diane's Office  
Source: City Water

## ANALYTICAL RESULTS

| Analyses | Method    | NELAP      |         | Qual | MRL      | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|-----------|---------------|---------|
|          |           | Accredited | Result  |      |          |       |           |               |         |
| Lead     | EPA 200.8 | A          | 0.00399 |      | 0.000103 | mg/L  | 0.020 AL  | 8/10/2016     | OML     |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

# Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

## Analysis Report

ORELAP 100016  
EPA OR00028

**Box R Waterlab**  
567 NW Second Street  
Prineville, OR 97754

Lab Order: **1608474**  
NRC Sample ID: **1608474-04A**  
Collection Date: **8/5/2016 8:15:00 AM**  
Received Date: **8/9/2016 10:00:00 AM**  
Reported Date: **8/16/2016 7:33:44 AM**

### Sample Information:

**X023294 L.J.H. School**

Client Sample ID: X023294-04  
Collectors Name: Wade Peasley  
Sample Location: Faucet on Top Floor  
Source: City Water

## ANALYTICAL RESULTS

| Analyses | Method    | NELAP      |         | Qual | MRL      | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|-----------|---------------|---------|
|          |           | Accredited | Result  |      |          |       |           |               |         |
| Lead     | EPA 200.8 | A          | 0.00190 |      | 0.000103 | mg/L  | 0.020 AL  | 8/10/2016     | OML     |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

CLIENT: Box R Waterlab  
 Work Order: 1608474  
 Project: X023294 L.J.H. School

**ANALYTICAL QC SUMMARY REPORT**

TestCode: ICPMS\_200.8\_SCHOOL

|            |          |           |       |           |             |             |      |                |           |           |             |      |          |      |
|------------|----------|-----------|-------|-----------|-------------|-------------|------|----------------|-----------|-----------|-------------|------|----------|------|
| Sample ID  | MB-36184 | SampType: | MBLK  | TestCode: | ICPMS_200.8 | Units:      | mg/L | Prep Date:     | 8/10/2016 | RunNo:    | 89136       |      |          |      |
| Client ID: | ZZZZZ    | Batch ID: | 36184 | TestNo:   | EPA 200.8   | (EPA 200.8) |      | Analysis Date: | 8/10/2016 | SeqNo:    | 1342686     |      |          |      |
| Analyte    |          | Result    |       | MRL       | SPK value   | SPK Ref Val |      | %REC           | LowLimit  | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |

|      |  |    |  |          |  |  |  |  |  |  |  |  |  |  |
|------|--|----|--|----------|--|--|--|--|--|--|--|--|--|--|
| Lead |  | ND |  | 0.000103 |  |  |  |  |  |  |  |  |  |  |
|------|--|----|--|----------|--|--|--|--|--|--|--|--|--|--|

|            |           |           |       |           |             |             |      |                |           |           |             |      |          |      |
|------------|-----------|-----------|-------|-----------|-------------|-------------|------|----------------|-----------|-----------|-------------|------|----------|------|
| Sample ID  | LCS-36184 | SampType: | LCS   | TestCode: | ICPMS_200.8 | Units:      | mg/L | Prep Date:     | 8/10/2016 | RunNo:    | 89136       |      |          |      |
| Client ID: | ZZZZZ     | Batch ID: | 36184 | TestNo:   | EPA 200.8   | (EPA 200.8) |      | Analysis Date: | 8/10/2016 | SeqNo:    | 1342687     |      |          |      |
| Analyte    |           | Result    |       | MRL       | SPK value   | SPK Ref Val |      | %REC           | LowLimit  | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |

|      |  |         |  |          |     |   |  |      |    |     |  |  |  |  |
|------|--|---------|--|----------|-----|---|--|------|----|-----|--|--|--|--|
| Lead |  | 0.09878 |  | 0.000104 | 0.1 | 0 |  | 98.8 | 85 | 115 |  |  |  |  |
|------|--|---------|--|----------|-----|---|--|------|----|-----|--|--|--|--|

|            |               |           |       |           |             |             |      |                |           |           |             |      |          |      |
|------------|---------------|-----------|-------|-----------|-------------|-------------|------|----------------|-----------|-----------|-------------|------|----------|------|
| Sample ID  | 1608520-49AMS | SampType: | MS    | TestCode: | ICPMS_200.8 | Units:      | mg/L | Prep Date:     | 8/10/2016 | RunNo:    | 89136       |      |          |      |
| Client ID: | ZZZZZ         | Batch ID: | 36184 | TestNo:   | EPA 200.8   | (EPA 200.8) |      | Analysis Date: | 8/10/2016 | SeqNo:    | 1342698     |      |          |      |
| Analyte    |               | Result    |       | MRL       | SPK value   | SPK Ref Val |      | %REC           | LowLimit  | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |

|      |  |         |  |          |     |            |  |      |    |     |  |  |  |  |
|------|--|---------|--|----------|-----|------------|--|------|----|-----|--|--|--|--|
| Lead |  | 0.09560 |  | 0.000104 | 0.1 | 0.00004406 |  | 95.6 | 70 | 130 |  |  |  |  |
|------|--|---------|--|----------|-----|------------|--|------|----|-----|--|--|--|--|

|            |                |           |       |           |             |             |      |                |           |           |             |      |          |      |
|------------|----------------|-----------|-------|-----------|-------------|-------------|------|----------------|-----------|-----------|-------------|------|----------|------|
| Sample ID  | 1608520-49AMSD | SampType: | MSD   | TestCode: | ICPMS_200.8 | Units:      | mg/L | Prep Date:     | 8/10/2016 | RunNo:    | 89136       |      |          |      |
| Client ID: | ZZZZZ          | Batch ID: | 36184 | TestNo:   | EPA 200.8   | (EPA 200.8) |      | Analysis Date: | 8/10/2016 | SeqNo:    | 1342699     |      |          |      |
| Analyte    |                | Result    |       | MRL       | SPK value   | SPK Ref Val |      | %REC           | LowLimit  | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |

|      |  |         |  |          |     |            |  |      |    |     |        |      |    |  |
|------|--|---------|--|----------|-----|------------|--|------|----|-----|--------|------|----|--|
| Lead |  | 0.09684 |  | 0.000104 | 0.1 | 0.00004406 |  | 96.8 | 70 | 130 | 0.0956 | 1.30 | 20 |  |
|------|--|---------|--|----------|-----|------------|--|------|----|-----|--------|------|----|--|

Qualifiers: E Value above quantitation range      H Holding times for preparation or analysis exceeded      J Analyte detected below quantitation limits  
 ND Not Detected at the Minimum Reporting Limit      R RPD outside accepted recovery limits      S Spike Recovery outside accepted recovery limits



